Response regarding an unusual shaped IUS

I read with interest Dr MacGregor's letter¹ regarding the unusual shape of a Mirena[®] intrauterine system (IUS) following removal. From the photograph it would appear that the capsule has become displaced and the arms of the device enclosed within it.

We have seen a case of complete detachment of the capsule from the frame, with unrecognised retention of the capsule within the uterine cavity. This was only recognised at hysteroscopy for investigation of subsequent infertility. This was described in a letter to the *British Journal of Obstetrics and Gynaecology* in 2008.²

It would seem prudent to check the integrity of the IUS at removal.

Adam Forrest, MBChB, MRCOG

Obstetrics and Gynaecology Registrar, Department of Obstetrics and Gynaecology, Derriford Hospital, Plymouth, UK; adam_forrest@btopenworld.com

Competing interests None.

Journal of Family Planning and Reproductive Health Care 2013;**39**:154.

doi:10.1136/jfprhc-2013-100610

REFERENCES

- 1 MacGregor A. An unusual shaped IUS. *J Fam Plann Reprod Health Care* 2013;**39**:64.
- 2 Forrest A, Amarakone I, Lord J. Retained hormone release capsule following removal of Mirena intrauterine system. *BJOG* 2008;**115**:130–131.

LETTERS TO THE EDITOR

Letters to the Editor are welcome and generally should not exceed 600 words or cite more than five references. For comments on material published in the most recent issue of the Journal, correspondence should be received within 4 weeks of dispatch of that Journal to be in time for inclusion in the next issue. When submitting letters correspondents should include their job title(s) and their work affiliation(s)/contact address(es). A statement on competing interests should also be submitted for all letters. Letters may be submitted to the Journal Editor or to the Journal Editorial Office (details on the Editorial Board page).