Comment on 'Safety issue with TT380 Slimline intrauterine contraceptive device'

- 2 Parker CG. Response to 'Safety issue with TT380 Slimline intrauterine contraceptive device'. *J Fam Plann Reprod Health Care* 2017;43:242.

I refer to the letter by Dr Trewinnard¹ in the July 2017 edition of this journal regarding safety concerns with the TT380 intrauterine device (IUD).

In my experience, the inserter tube of the TT380 Slimline device is not malleable enough to shape into a semicircular form for insertion of the IUD into a retroverted or acutely anteverted uterus. Attempts to persist with shaping the inserter will result in kinking of the inserter tube with occlusion of its lumen and perforation of the tube by the introducer rod if insertion is attempted, as experienced by Dr Trewinnard.

Failure to shape the inserter tube will thus result in a failed IUD insertion in the aforementioned uterine positions and worse, possible uterine perforation.

I have stopped using the TT380 Slimline for the past 6 years because of this problem, and now only use the NovaT 380 (Bayer Pharma AG) and T-Safe 380 (Mona Lisa N.V.) systems.

I look forward to the improved redesigned TT380 Slimline.²

Disraeli Okoro

Correspondence to Dr Disraeli Okoro, Camberwell Sexual Health Clinic, King's College NHS Trust Hospital, London, UK; d.okoro@nhs.net

Competing interests None declared.

Provenance and peer review Not commissioned; internally peer reviewed.

© Article author(s) (or their employer(s) unless otherwise stated in the text of the article) 2018. All rights reserved. No commercial use is permitted unless otherwise expressly granted.

CrossMark

Published Online First 8 November 2017

BMJ Sex Reprod Health 2018;44:68.
doi:10.1136/jfprhc-2017-101871

REFERENCES

- 1 Trewinnard K. Safety issue with TT380 Slimline intrauterine contraceptive device. *J Fam Plann Reprod Health Care* 2017;43:241–242.